

Brihanmumbai Mahanagarpalika
TNMC and B.Y.L. Nair Ch. Hospital, Mumbai
Department of Paediatric Surgery

Outward Number/ Tender Number:

Date: 23/07/2024

M/s.....

Subject: Procurement of Teaching aids/ Medicines/Surgical items/Equipments/Other approved items through DDF funds.


Sir,

Sealed quotations are invited for the attached listed items.

1. Name/Names of the item
2. Rate: basic rate + GST =total
3. Vendor Name:
4. Name of the Proprietor:
5. Contact number of Authorised person:
6. Company Stamp and valid signature.
7. **Technical Bid & commercial bid in separate packets (Two separately marked and sealed envelopes)**

Envelop to be addressed to	Dr Hemanshi Shah, Professor and Head, Department of Paediatric Surgery
Start Date to submit the Sealed Quotation	25/07/2024
Last date to submit the quotation	10/08/2024
Place of Submission of sealed envelops	DDF office, First Floor, G-Building
Time for submission	10.30 am to 1.30 pm

Terms and conditions and other relevant details are mentioned in Bid document attached.



Professor and Head
Department of Paediatric Surgery
T.N. Medical College &
B.Y.L. Nair Ch. Hospital, Mumbai

Sr. No.	Name of Department	BID No	Short title of Item to be Purchased	Start Date for BID Date & Time	End Date for BID Date & Time on /Before 10 th August at 10: 30 to 1:00pm	Std BID Document	Remarks
1	Dept of Paediatric surgery	Paed Surg 17 Dt. 23/7/2024	<ol style="list-style-type: none"> 1. Paediatric Compact Approx. 2.5Fr / 3.8Fr Cystourethroscope with working channel with accessories 2. Laparoscopic simulator endotrainer with reusable surgical instruments and monitor 3. Fine specialised Neonatal Steel Surgical Instruments (List attached) 4. Digital baby & infant weighing scale 5. Stainless steel cidex tray with cover 6. Lightweight transparent formalin chamber 	25/07/2024 10:30 am	10/08/2024 10:30-1:00 pm	Specifications and Brochures	

Hemanshi Shah
Dr. HEMANSHI SHAH
Professor & Head,
Dept. of Pediatric Surgery
T. N. Medical College &
B.Y.L. Nair Ch. Hospital
Mumbai - 400008.

Sr No	Name of Equipment	Specifications	No of Required	Special Instruction
1	Paediatric Compact Approx. 8Fr/9.8Fr Cystourethroscope with working channel with accessories (grasping forcep , bugbee electrode , reprocessing basket, etc)	<ul style="list-style-type: none"> • It should be approx 8/9.8 Fr Diameter, angled/straight ocular with length of approx. 160mm • With approx 7 degree direction of View, Autoclavable • It should have more than 4.0 Fr working channel • It should be supplied with autoclaveable instrument tray • It should have straight channel • It shuld have maintenance free stopcocks • It should have oval channel design with excellent flow characteristics • It should have ergonomic design • It should be made up of stainless steel 	1	<ul style="list-style-type: none"> • Should provide 1 year warranty • It should follow international standards of safety requirements European CE/FDA certified • Demonstration compulsory • Training to MEC engineers • Providing operational and service manual • Power supply of 220 +/- 15V, 50Hz
2	Laparoscopic simulator endotrainer with reusable surgical instruments – (Maryland , grasper , needle holder , scissor)and monitor	<ul style="list-style-type: none"> • dome shape • inbuilt camera and light source • dimension approx. 450mmx190mm x 295mm • compatible converter cable 	1	<ul style="list-style-type: none"> • Should provide 1 year warranty • It should follow international standards of safety requirements European CE/FDA certified • Demonstration compulsory • Training to MEC engineers • Providing operational and service manual • power supply 100-240w AC frequency- 50-60 Hz
3	Fine specialised Neonatal Steel Surgical Instruments	<ul style="list-style-type: none"> • FINE ADSONS TOOTH 5" • FINE ADSONS PLAIN 5" • ADSONS TOOTH 5" • ADDSONS pLAIN 5" • PLAIN THUMB FORCEPS • MCINDOE FORCEP 	<p>4</p> <p>4</p> <p>2</p> <p>2</p> <p>2</p> <p>2</p>	<ul style="list-style-type: none"> • Made of Stain less steel • Autoclavable • Size and fineness as per the list • Reuseable • 1 year warranty • ISO Certified
		<ul style="list-style-type: none"> • MASTOID RETRACTOR SHARP • MOLLISONS MASTOID RETRACTOR 	<p>2</p> <p>2</p>	
		<ul style="list-style-type: none"> • DOLPHIN SCISSORS 4" 	2	
		<ul style="list-style-type: none"> • BABY STILLEY SCISSOR 4" • STILLEY SCISSOR 6" 	<p>2</p> <p>2</p>	

		<ul style="list-style-type: none"> • SUTURE CUTTING SCISSOR 4" • BABY MOSQUITO 4" • SMALL STRAIGHT ARTERY 5" 	1 6 6	
		<ul style="list-style-type: none"> • NEEDLE HOLDER SMALL FINE 5" • NEEDLE HOLDER BIG 7" • BABY MIXTER 5" • MIXTER 6" 	2 2 2 2	
		<ul style="list-style-type: none"> • BABY BABOCK 4" • BABY ALLIS 4" 	2 3	
		<ul style="list-style-type: none"> • FINO CHIETTO • NEONATAL DEAVERS • BULL DOG CLAMP small 	1 2 2	
4	Digital baby & infant weighing scale	<ul style="list-style-type: none"> • Portable • Weighing capacity Approximately Up to 20 Kg • Along with converting power cable • Large LCD Display • Auto on and Off 	2	<ul style="list-style-type: none"> • ISO Certified • 1 year warranty • Demonstration compulsory • Training to MEC enginers • Providing operational and service manual • Power supply of 220 +/- 15V, 50Hz
5	Stainless steel cidex tray with cover	<ul style="list-style-type: none"> • Portable • Stainless steel • Auto clavable • Non Corrosive • Leak proof • Dimension Approximately 74 cm x 14 cm x 20 cm 	2	N.A
6	Lightweight transparent formalin chamber	<ul style="list-style-type: none"> • Dimension Approximately 66 cm x 21 cm x 21 cm • Portable • Non Corrosive • Leak proof • Transparent, Fiber • 3 Shelves 	1	N.A

Hemanshi Shah

Dr.Hemanshi shah
Professor & Head
Dept. of Pediatric Surgery