

Brihanmumbai Mahanagarpalika
TNMC and B.Y.L. Nair Ch. Hospital, Mumbai
Department of Nephrology

Outward Number/ Tender Number: NEPH-287

Date: 29.07.2024

M/s.....

Subject: Procurement of Electrohydraulic Operation Table (1 no) through DDF funds.

Sir,

Sealed quotations are invited for the attached listed items.

1. Name/Names of the item:
2. Rate: basic rate + GST = total
3. Vendor Name:
4. Name of the Proprietor:
5. Contact number of Authorised person:
6. Company Stamp and valid signature.
7. Technical Bid & commercial bid in separate packets

Envelop to be addressed to	Professor and Head, Department name
Start Date to submit the Sealed Quotation	01/08/2024
Last date to submit the quotation	10/08/2024
Place of Submission of sealed envelops	DDF office, First Floor, G-Building
Time for submission	10.30 am to 1.30 pm

Terms and conditions and other relevant details are mentioned in Bid document attached.

[Handwritten Signature]
31/07/2024

Professor and Head
Department of Nephrology
T.N. Medical College &
B.Y.L. Nair Ch. Hospital, Mumbai

Format(Prepare in landscape)

Vendor Name:

Name of the Proprietor:

Contact Name & Number of Authorised Person:

S. No.	Name of the Item	Company & Specification	Warranty/ AMC Specification	Warranty/ AMC Details	Rate+ GST/ Item	Total Numbers	Final Amount

Company stamp and valid signature

S.NO.	Item name	Specification	Requirement	Special Instructions
1	Electrohydraulic Operation Table	<ul style="list-style-type: none"> • Operation table with Electro Hydraulic System • Fully operated by electrical remote with dual console operating separately. One control on the body and one hand held with flexible cord. • Movements : Up down, lateral tilt, Trendelenburg / Reverse position + Degree / 20 degree. • C-arm compatible, Top sliding, Moving on either side completely (300 mm on either side) • Glassfiber / Helium laminated, scratch proof • Length 1900 mm to 2100 mm • Width 500 mm to 600 mm • Accessories – Arm rests, Horizontal extension with support for hand surgeries. • Full Auto zero leveling switch. • Provided with anti-bacterial molded foam mattress. (Certificate to be attached) • Minimum height of table should be 500 mm, maximum height 1200 mm (Adjustable) • Wight bearing capacity should be between 150 kg static & dynamic. • Floor locking mechanism to be operated electrically through Remote Control only. Foot peddle locking will not be accepted. 	1 nos	<ul style="list-style-type: none"> • Delivery should be within 30 days of Purchase order. • 04 Maintenance service should be done yearly. • Breakdown call should be attended within 24 hours.

		<ul style="list-style-type: none">• Table should have a inbuilt battery backup for emergency operation in case of power failure. The battery should last for at least 50 operations of the table.• Preferably US FDA (Approval Certificate with K number documents) / European CE from Notified body / ISO 13485 approved produced. Non notified body CE certificate will not be accepted. Electrical safety conforms to the standards for electrical safety IEC 60601 requirements.• Should have 03 years warranty• Should have 05 years CMC after warranty period get over. The total cost of the machine must include CMC cost as well.	
--	--	---	--

Signature
31/07/2024

**Professor and Head
Department of Nephrology
T.N. Medical College &
B.Y.L. Nair Ch. Hospital, Mumbai**